



# Specification

---

1.09468.0000 Buffer tablets pH 7.2 for preparing buffer solution acc. to WEISE for staining of blood smears

---

---

## Specification

---

pH-value (1 tablet in 1 l of water)	7.15 - 7.25	
Crushing strength	≥ 50	N
Disintegration time	≤ 20	min
Suitability for blood-smear staining	conforms	

Dr. Timm Schlegelmilch

---

Responsible laboratory manager quality control

This document has been produced electronically and is valid without a signature.