

Technical Note

Demonstrating Millipore Express® SHR mycoplasma clearance capability at pilot and production scale over an extended filtration time

Introduction

Two separate *Acholeplasma laidlawii* retention studies were performed using sterilizing-grade Millipore Express® SHR with Prefilter (0.5/0.1 µm polyethersulfone) filters and *Acholeplasma laidlawii* ATCC® 23206™ to determine their mycoplasma retentive properties over an extended filtration time. In these experiments, the conditions of the test were modeled after typical processing parameters and were conducted using a low organism challenge level and typical operating pressures. The first experiment tested filters at production scale in a 3 x 30" cartridge configuration. The filters were fully retentive when challenged with 200 L of *A. laidlawii* at a level of 4.85×10^4 CFU/cm² over five hours. A second experiment tested the retention of *A. laidlawii* using Opticap® XL 300 pilot scale capsules. The Opticap® XL 300 capsules were fully retentive after an *A. laidlawii* challenge of 7.81×10^5 CFU/cm² for eight hours.

Background

Currently, no standard industry method exists for testing filters for mycoplasma retention; however, Merck Millipore has a validated procedure for mycoplasma testing that is modeled after the current ASTM® F838-05(2013) test for sterilizing-grade filters.

Similar to the ASTM® test protocol, Merck Millipore's validated test method involves using a high challenge level of *A. laidlawii* ($\geq 10^7$ CFU/cm² effective filtration area, EFA) and a high differential pressure (30 psid). These severe testing parameters are used to simulate worst case testing parameters and provide a high degree of assurance that

a filter that passes this test will quantitatively retain large numbers of organisms².

While this testing regime is useful in providing a quantitative assessment of sterilizing-grade filter performance, these severe testing conditions are not typical of a pharmaceutical biomanufacturing process. As a result, the following two experiments were designed to examine filter performance at pilot and production scale under conditions which more closely models typical cell culture filtration operating parameters.

Procedure

Materials And Equipment

Production Scale Testing: 3 x 30" Configuration

The 3 x 30" filter extended challenge test was scaled from an established filtration process¹ using 47 mm membrane disc filters (13.8 cm² of total effective filtration area). For this test, three 30-inch Millipore Express[®] SHR with Prefilter cartridges (44,100 cm² of total effective filtration area) were challenged over five hours at a flow rate of 667 mL/min (flux = 9.1 LMH). Following the filter challenge, *A. laidlawii* culturing, recovery and enumeration, were performed.

Pilot Scale Testing: Opticap[®] XL 300 Configuration

The Opticap[®] XL 300 (290 cm² effective filtration area) small scale capsule (SSC) extended challenge test¹ was adapted with a few modifications including: test stand setup, challenge and flush volumes. *A. laidlawii* culturing, recovery and enumeration was also performed³. For the duration of the test, pressure was driven at 10 psid with a flow of 300 mL/min (flux = 620.7 LMH) for an extended challenge time of eight hours.

Membrane	Millipore Express [®] SHR with Prefilter 0.5/0.1 µm	
Test filter	3 x 30" code 7 cartridges	Opticap [®] XL 300 capsule
EFA (cm ²)	44,100	290
Equipment	<ul style="list-style-type: none">• Mobius[®] MIX200 disposable mixing assembly• 3 x 30" multi-round code 7 cartridge housing• Sterifil[®] aseptic system• Aervent[®]-50 vent filter• Assay filter: Durapore[®] 0.22 µm PVDF 47 mm analytical membrane discs• Gauges: sanitized and calibrated• Pump: peristaltic• Tubing: silicone	<ul style="list-style-type: none">• Pressure pots• Sterifil[®] aseptic system• Aervent[®]-50 vent filter• Assay filter: Durapore[®] 0.22 µm PVDF 47 mm analytical membrane discs• Gauges: sanitized and calibrated• Pump: peristaltic• Tubing: silicone

Fatty Acid Supplements

Oleic and palmitic acids were prepared separately in 100% ethanol to achieve a concentration of 10 mg/mL each. Each fatty acid was sterile filtered and stored separately.

Glucose Hydrolysate Broth (GHB)

Sterile GHB contains: 4% w/v polypeptone, 0.5% w/v Trizma[®] base (2-amino-2-(hydroxymethyl)-1,3-propanediol), 0.78% glucose, and 0.4% bovine serum albumin (BSA) in Milli-Q[®] water. Oleic acid and palmitic acid were aseptically added to deliver a final concentration of 0.002% w/v each.

Glucose Mycoplasma Agar (GMA)

Sterile GMA contains: 2% w/v Mycoplasma Broth Base, 0.5% w/v Trizma[®] base, 0.78% w/v glucose, 0.4% w/v BSA and ASTM[®] Type 1 water. The agar base was prepared by resuspending 1.2 % w/v agar in Milli-Q[®] water and

sterilizing in an autoclave with a validated slow exhaust cycle at a minimum of 121.1 °C. The agar base was tempered and aseptically added to the sterile GMA to arrive at a dispensing temperature of 43 to 45 °C. Oleic acid and palmitic acid were aseptically added to deliver a final concentration of 0.002% w/v each. Also, 2,3,4-triphenyltetrazolium chloride (TTC) was added aseptically for a final concentration of 0.05% v/v.

Diluent and Resuspension Buffer

Mycoplasma phosphate buffer: sodium phosphate monobasic, NaH₂PO₄, 28 mM and sodium phosphate dibasic, Na₂HPO₄, 72 mM were dissolved in Type 1 water and adjusted to pH 7.1 (± 0.2). The buffer was sterilized by autoclaving in a validated slow exhaust cycle at a minimum of 121 °C.

Procedure

Test Equipment Preparation

Production Scale Testing: 3 x 30" Configuration

The test cartridges (three 30" cartridges) were installed into a multi-round filter housing and the cartridges and test system were steam sterilized. After steaming, the test system was cooled using filter sterilized compressed air.

Analytical membrane filters (Durapore® 0.22 µm) were installed into Sterifil® aseptic systems. The Sterifil® aseptic systems, challenge test manifolds, and other associated equipment were wrapped and sterilized in a validated autoclave at a minimum of 121 °C.

Pilot Scale Testing: Opticap® XL 300 Configuration

The Opticap® XL 300 capsule was attached to stainless steel fittings and the entire assembly was sterilized via autoclave, using a validated liquid cycle at 121 °C for 20 minutes⁴.

Analytical membrane filters (Durapore® 0.22 µm) were installed into Sterifil® aseptic systems. The Sterifil® aseptic systems, challenge test manifolds, and other associated equipment were wrapped and sterilized in a validated autoclave cycle at a minimum of 121 °C.

Culture Preparation

A frozen stock vial of *A. laidlawii* was thawed rapidly and transferred to GHB broth to achieve a final concentration of 4% (v/v) inoculum to broth. Cultures were incubated at 37 °C (± 2 °C), with 6% (± 1%) CO₂ for 22 (± 2) hours.

Challenge Suspension Preparation

3 x 30" Configuration

A. laidlawii from a 22 (± 2) hour culture in GHB was diluted into 200 L of mycoplasma phosphate buffer to deliver a final target concentration of 10⁴ CFU/cm² of membrane effective frontal surface area. The cartridges provided an effective frontal area (EFA) of 44,100 cm².

Opticap® XL 300

A. laidlawii from a 22 (± 2) hour culture in GHB was diluted into 2 L of mycoplasma phosphate buffer to deliver a final target concentration of 10⁴ CFU/cm² of membrane effective frontal surface area. The Opticap® XL 300 capsule provided an EFA of 290 cm².

Sterility Test

The test system sterility was assessed by flowing sterile mycoplasma phosphate buffer through the test system. For the 3 x 30" configuration, a 1 L filtrate sample was taken. For the Opticap® XL 300 capsule, a 2 L sample was taken.

Bacterial Challenge Test

The 3 x 30" cartridge retention test was a dead-end configuration. The *A. laidlawii* challenge suspension was delivered to the challenge filters by a peristaltic pump with a constant flow of 667 mL/min (flux = 9.1 LMH) and the filtrate was collected in a 200 L Mobius® collection bag. Once every hour, a 1 L sample was taken from a sample port installed before the filtrate collection bag, starting after one hour of test time, and ending after five hours for a total of five samples.

For the Opticap® XL 300 capsule experiment the 2 L challenge was delivered via peristaltic pump and the entire challenge was collected downstream and assayed by membrane filtration. After the dead-end challenge portion, the Opticap® XL 300 capsule experiment was conducted in a recirculation configuration, in which 2000 mL of mycoplasma buffer in a separate sterile flask was recirculated through the Opticap® XL 300 Capsule at a rate of 300 mL/min (flux = 620.7 LMH) for eight hours. After eight hours, the 2 L filtrate was then assayed by membrane filtration.

Challenge Titer

3 x 30" Cartridge Configuration

Samples of the *A. laidlawii* culture, initial *A. laidlawii* challenge (titer taken from the Mobius® mixing system), and intermittent samples (taken upstream of the test filters), were serially diluted in mycoplasma buffer and enumerated using the pour-plate method with GMA agar³. Plates were incubated at 37 °C (± 2 °C), with 6% (± 1%) CO₂ for four days. Post-incubation, the pour-plates were enumerated and the titer was determined.

Opticap® XL 300

A sample from the challenge flask was taken immediately prior to testing and assayed by the pour plate method with GMA agar³. Plates were incubated at 37 °C (± 2 °C), with 6% (± 1%) CO₂ for four days. Post-incubation the pour-plates were enumerated and the titer was determined.




Figure 1. Process flow diagram for the 3 x 30 inch challenge.

Recovery Assay

The Membrane Filtration (MF) method was employed for enumeration of colonies in the filtrate. For both experiments, the effluent was filtered through a Sterifil® system containing a Durapore® 0.22 µm analytical

membrane. Each analytical membrane was then aseptically transferred to a separate GMA agar plate and incubated at 37 °C (± 2 °C), with 6% (± 1%) CO₂ for four days, to test mycoplasma retention.

Calculations

The pour-plates were used to calculate the bacterial challenge concentration in CFU/filter and the EFA challenge in CFU/cm². The MF-Millipore™ filter plates were used to determine system sterility and test filter passage.

Equation 1 Total challenge in CFU/filter

$$\text{Total challenge} = \text{challenge concentration (CFU/mL)} \times \text{challenge volume (mL/filter)}$$

Equation 2 EFA (effective frontal area) challenge in CFU/cm²

$$\text{EFA challenge} = \frac{\text{total challenge (CFU/filter)}}{\text{filter surface area (cm}^2\text{/filter)}}$$

Equation 3 Total passage in CFU/filter

$$\text{Total passage} = \text{total CFU on the analytical membrane}$$

Equation 4 Log Reduction Value

$$\text{LRV} = \log_{10} \left(\frac{\text{total challenge (CFU/Filter)}}{\text{total passage (CFU/Filter)}} \right)$$

If the total passage for a sample is zero (absence of growth), the raw data is recorded as 0 CFU. However, to calculate LRV a "1" is placed in the denominator to represent the Total Passage and the LRV is reported with a "greater than" symbol.

Equation 5 Filtrate flux (LMH)

$$\text{Flux} = \frac{\text{filtrate flow rate (mL/min)/membrane area (cm}^2\text{)} \times 600}{\text{cm}^2 \times 600}$$

Figure 2. Representative equations

Acceptance Criteria

System Sterility: Absence of microbial growth.

Results

For all experiments, the system sterility met the acceptance criteria demonstrating the absence of microbial growth and indicating that all equipment was properly sterilized. The average delivered challenge level was 4.85×10^4 CFU/cm² for the 3 x 30" configuration and 7.81×10^5 CFU/cm² for the Opticap® XL 300 capsules. There was zero passage of challenge organism detected downstream of the 3 x 30" cartridge filters (LRV ≥ 4.0). The limit of detection for the 3 x 30" configuration experiment was 40 CFU. There was zero passage of challenge organism detected downstream of the Opticap® XL 300 capsules (LRV ≥ 8.35); the limit of detection for these experiments was 1 CFU.

Conclusion

Challenge testing performed at pilot and production scale using *A. laidlawii* demonstrated that Millipore Express® SHR with Prefilter cartridge filters in a 3 x 30" configuration were fully retentive when challenged with 200 liters of *A. laidlawii* at a concentration of 4.85×10^4 CFU/cm² over five hours. The Opticap® XL 300 small scale capsules were also fully retentive when challenged at an average *A. laidlawii* level of 7.81×10^5 CFU/cm² over eight hours. These pilot and production scale filtration tests demonstrate that Millipore Express® SHR filter mycoplasma retention capability at typical cell culture filtration operating parameters and extended processing times. Under these conditions, Millipore Express® SHR filters are retentive for relative concentrations and volumes.

References

1. *Acholeplasma laidlawii* Retention Test of Flatstock Membrane (Merck Millipore Document #0003007TM)
2. American Society for Testing and Materials. Standard Method for Determining Bacterial Retention of Membrane Filters Utilized for Liquid Filtration. 2005 ASTM® Standards on Materials and Environmental Microbiology, Second Edition, Designation F838-05(2013).
3. Merck Millipore document #00081823SO "Culturing, Enumeration and Challenge Preparation of *Acholeplasma laidlawii*."
4. Merck Millipore document #065473SO ("Getting Autoclave Model 91415 Validated Load Profile & Operator").

To Place an Order or Receive Technical Assistance

In Europe, please call Customer Service:

France: 0825 045 645

Germany: 01805 045 645

Italy: 848 845 645

Spain: 901 516 645 Option 1

Switzerland: 0848 645 645

United Kingdom: 0870 900 4645

For other countries across Europe,
please call: +44 (0) 115 943 0840

Or visit: www.merckmillipore.com/offices

For Technical Service visit:

www.merckmillipore.com/techservice



www.merckmillipore.com/offices