3050 Spruce Street, St. Louis, MO 63103 USA
Tel: (800) 521-8956 (314) 771-5765 Fax: (800) 325-5052 (314) 771-5757
email: techservice@sial.com sigma-aldrich.com

Product Information

Cannabidiol solution
1.0 mg/mL±5% in methanol
analytical standard for drug analysis

Catalog Number **C6395** Lot Number SLBM6755V Storage Temperature 2–8 °C

CAS RN 13956-29-1

Synonyms: CBD, 2-[3-Methyl-6-(1-methylethenyl)-2-cyclohexen-1yl]-5-pentyl-1,3-benzenediol

$$H_3C$$
 HO $CH_2(CH_2)_3CH_3$ H_2C CH_3

Product Description

Molecular Formula: C₂₁H₃₀O₂ Molecular Weight: 314.46

DEA Class: Exempt preparation of Class I

Concentration: 1.00 mg of cannabidiol/mL of GC grade methanol, concentration verified by UV.

Expiration Date: March 2018

Precautions and Disclaimer


This product is for R&D use only, not for drug, household, or other uses. Please consult the Safety Data Sheet for information regarding hazards and safe handling practices.

Storage/Stability

Store the product at 2–8 °C. If stored at 2–8 °C in an airtight container, protected from light, decomposition of the product is less than 1% in 48 months. After opening, concentration may change due to loss of solvent.


CMW,MAM 04/15-1

70 eV Electron Ionization Mass Spectrum



Hewlett-Packard 6890 Series II Gas Chromatograph Hewlett-Packard 5973 Mass Selective Detector Hewlett-Packard G1701DA MS ChemStation (HP-DOS series)

FT-IR





Ultraviolet Spectrum

Capillary Gas Chromatography

Peak	273.5 nm
Absorbance	0.3466
Solvent	GC grade methanol
Dilution	1:9 (v/v)

Column Supelco SPB-1 (Catalog Number 24162)
I.D. 0.20 mm Length 15 m Film Thickness 0.20 μm
Oven Temperature 150 °C (2 minutes), then
15 °C/minute to 300 °C, hold 11 minutes
Injector Temperature 250 °C, Split
Detector Hewlett-Packard 5973
Mass Selective Detector (70 eV, 40-500 m/z)





Beckman-Coulter DU 800

Hewlett-Packard 6890