

Product Information

Holmium wavelength calibration standard

Catalog Number **H8015**
Lot Number 019K8713
Store at Room Temperature

Product Description

The Holmium wavelength calibration standard is a high quality secondary reference standard offered as a suitable alternative to the NBS Holmium Oxide Solution (No. 2034). It is used to establish the accuracy of conventional spectrophotometers in the spectral range 240 to 650 nm.

The product consists of holmium perchlorate (15% w/v) in 10% perchloric acid supplied as a 3 ml sample in a flame sealed glass ampule. The absorbance of the standard has been normalized against NBS Holmium Oxide Solution (No. 2034). The absorbance spectrum of Lot Number 019K8713 is shown (see Figure 1) along with the comparison of selected wavelength maxima of this lot versus the NBS Standard (see Table 1).

Precautions and Disclaimer

This product is for R&D use only, not for drug, household, or other uses. Please consult the Material Safety Data Sheet for information regarding hazards and safe handling practices.

Table 1.

Wavelength of Absorbance Maxima >0.2

H8015 Lot Number 019K8713	NBS Standard
241.5	241.5
277.5	277.5
286.5	286.5
360.5	360.5
416.0	416.0
450.5	450.5
484.5	484.5
536.0	536.0
640.0	640.0

Storage/Stability

Store the product at room temperature.


References

1. Graselli, J.G., CRC Atlas of Spectral Data and Physical Constants for Organic Compounds, (1973) pp. A-60.

CC,MAM 04/09-1

Figure 1.

Absorbance spectrum of Holmium perchlorate (15% w/v) in 10% perchloric acid (H8015, Lot Number 019K8713)



Conditions: 1 nm bandwidth; 0.5 nm data interval; 600 nm/minute scan speed; 1 cm pathlength; quartz cuvette
Instrument: Beckman-Coulter DU 800.

Sigma brand products are sold through Sigma-Aldrich, Inc.

Sigma-Aldrich, Inc. warrants that its products conform to the information contained in this and other Sigma-Aldrich publications. Purchaser must determine the suitability of the product(s) for their particular use. Additional terms and conditions may apply. Please see reverse side of the invoice or packing slip.