

QUALITY CONTROL RANGES



MILLIPLEX[®] MAP Rat/Mouse Neuropeptide Magnetic Bead Panel
Kit Catalog # RMNPMAG-83K
Control Catalog # RMNP-6083
Lot# RMNPM-104A and RMNPM-204A

<u>Analyte</u>	<u>QC Level</u>	<u>Expected Range</u>	<u>Units</u>
α -MSH	Control I	307 – 724	pg/mL
	Control II	1948 – 4047	pg/mL
β -Endorphin	Control I	606 – 1258	pg/mL
	Control II	3357 – 6972	pg/mL
Neurotensin	Control I	217 – 488	pg/mL
	Control II	1279 – 2656	pg/mL
Orexin A*	Control I	N/A	pg/mL
	Control II	N/A	pg/mL
Oxytocin	Control I	114 – 263	pg/mL
	Control II	686 – 1424	pg/mL
Substance P	Control I	22 – 50	pg/mL
	Control II	135 – 280	pg/mL

*** Control ranges are not available for this analyte.**

RMNP-6083: RMNPM-104A.204A (02 August 2019)
EMD Millipore • 1-800-645-5476 • 1-800-645-5439 (FAX) • emdmillipore.com